

ASSAY NAME: GIPARAA_BR384
(GI Parasite Panel A for Bio-Rad 384 or 96)

Quantity: 100 x 20µL PCR reactions

4-plex assay: *Cryptosporidium* spp., *Entamoeba histolytica*, *Giardia lamblia*, and human RPP30 DNA

SKU #: BUN-GIPARAA-D-BR384-100 (Bio-Rad)

(RUO). Research Use Only. Not for use in Diagnostic Procedures.

SCOPE OF THIS PRODUCT INFORMATION SHEET (PIS):

The oligonucleotide recipes are optimized for each instrument (QuantStudio, Bio-Rad). The verification data presented in this PIS were performed using BUN-GIPARAA-D-BR384-100 on a Bio-Rad CFX96 Real-Time System using white-bottomed plates. This kit is compatible with both CFX-96 and CFX-384 instruments. The performance of the other SKUs on their corresponding instrument should be similar. Contact PCRassays.com if you are planning to use a different instrument.

CONTENTS

The primers and probes in the GIPARAA_BR384 assay are provided in Tube 1 as a 5X concentrated working solution that detects 3 pathogens and a human control. The primers and probe for the *Cryptosporidium* spp. assay detects the following species: *C. parvum* and *C. hominis*. *Giardia lamblia* is also commonly known as *G. intestinalis* and *G. duodenalis* (modern name).

Table of Dyes used in this assay:

Pathogen/Target	Dyes	Quencher	Refs.
<i>Cryptosporidium</i> spp.	FAM	BHQ-1	1,2
RPP30-DNA control	HEX	BHQ-1	3
<i>E. histolytica</i>	TEX615	BHQ-2	4,5
<i>G. lamblia</i>	Cy5	BHQ-2	6

The probes are designed as a TaqMan⁷ cleavage mechanism and thus the reaction requires a DNA polymerase with 5'-exonuclease activity.

ASSAY HANDLING

The GIPARAA_BR384 bundle is shipped on ice and should be stored at -20 °C. The bundle should be kept on ice once thawed. Do not subject the enzyme to multiple freeze-thaw cycles.

Contamination should be avoided by using appropriate personal protective equipment (PPE), powder free gloves, aerosol barrier pipette tips, and a clean hood.

ASSAY CONTENTS:

Tube 1: 5X Primer/Probe mix for *Cryptosporidium* spp., *E. histolytica*, *G. lamblia*, and hRPP30DNA.

Tube 2: (Do NOT add to specimen unknowns) Positive control: Synthetic 500 bp DNA fragments for *Cryptosporidium* spp., *E. histolytica*, *G. lamblia*, and hRPP30DNA.

Tube 3: InhibiTaq qPCR enzyme Mastermix (enough for 100 rxns. with 20 µL total volume). This is a custom formulation from Fortis Life Sciences to the specifications of PCRassays.com.



EXPERIMENTAL

Perform nucleic acid extraction/purification (recommended).

Set up your PCR reactions (20 µL) as follows on ice:

Component	Volume (µL)
InhibiTaq enzyme mastermix (2X)	10
Primer/Probe mix (5X)	4
Sample	2
Water (Molecular Biology Grade)	4

Notes: To improve assay sensitivity, up to 6 µL of sample can be added (water volume adjusted accordingly) for a total reaction volume of 20 µL.

For positive control rxns., add 2 µL of the solution from Tube 2. For 10 µL reactions, divide all amounts above by a factor of 2.

A PCR protocol was used for verification on a Bio-Rad CFX96 System using white-bottomed plates, with the following program:

Step	Thermocycling Protocol:
1	Incubate @ 95 °C for 2 minutes
2	Incubate @ 95 °C for 3 seconds
3	Incubate @ 55 °C for 20 seconds*
4	Plate Read
5	Go to Step 2, repeat 44× more

* A 15 second extension time may be used for CFX96 instruments.

RESULT INTERPRETATION

After running the qPCR reaction, use the instrument software to determine the quantification cycle, C_q (or use C_T). Fluorescence channels with a $C_q < 38$ cycles, and final RFU > Threshold is considered “positive” or “+” in the Table below. The “Threshold” value for calling a PCR positive is dependent on the instrument model, well size, and sample volume; thus the user must determine the threshold that is appropriate for their method. For our Bio-Rad CFX96 with 200 µL wells and 20 µL reaction volume, the average RFU was approximately 2,200 and we used a threshold of 220 for calling positives or “+” in the Table below.

<i>Cryptosporidium</i> spp. FAM™	hRPP30 HEX™	<i>E. histolytica</i> TEX615™	<i>G. lamblia</i> Cy5™	Recommended Interpretation
—	—	—	—	The PCR reaction failed. Please repeat the experiment.
—	+	—	—	The sample does not contain pathogen DNA. The sample contains human RPP30 DNA.
+	—	—	—	The sample contains <i>Cryptosporidium</i> spp. DNA. The sample may not contain human RPP30 DNA.
+	+	—	—	The sample contains <i>Cryptosporidium</i> spp. DNA and human RPP30 DNA.
—	—	+	—	The sample contains <i>E. histolytica</i> DNA. The sample may not contain human RPP30 DNA.
—	+	+	—	The sample contains <i>E. histolytica</i> DNA and human RPP30 DNA.
—	—	—	+	The sample contains <i>G. lamblia</i> DNA. The sample may not contain human RPP30 DNA.
—	+	—	+	The sample contains <i>G. lamblia</i> DNA and human RPP30 DNA.
+	—	+	+	The sample contains <i>Cryptosporidium</i> spp., <i>E. histolytica</i> , <i>G. lamblia</i> DNA. The sample may not contain human RPP30 DNA.
+	+	+	+	The sample contains <i>Cryptosporidium</i> spp., <i>E. histolytica</i> , <i>G. lamblia</i> DNA and human RPP30 DNA.

VERIFICATION EXPERIMENTS

The GIPARAA_BR384 assay verification was carried out as a 4-plex assay, which simultaneously detects DNA from *Cryptosporidium* spp., *E. histolytica*, *G. lamblia*, and human RPP30 DNA, which serves as a positive extraction-control assay.

Experiments were performed in triplicate using the experimental procedure given above, but with different samples added to each reaction. The samples used for the verification experiments contained 1×10^4 copies/reaction of 500 bp synthetic DNA constructs (from Twist Biosciences) harboring the regions of interest from the pathogen genomes, human RPP30 DNA gene, and human genomic DNA. The results of these experiments are shown in **Figure 1** and indicate that the 4-plex specifically detects the different pathogens in the human genomic DNA matrix.

The limit of detection (LOD) was estimated by performing serial dilution experiments in triplicate (**Figure 2**). For dilution series only target construct was added. The results show a limit of detection (LOD) <10 copies/reaction.

Conclusion: The data in **Figure 1** indicate that the 4-plex primers and probes specifically detect and differentiate the pathogens and are also compatible with RPP30_DNA positive control primers.

NOTES

¹ FAM™ (Carboxyfluorescein) is a trademark of Life Technologies, Inc

² BHQ-1™ (Black Hole Quencher) is a trademark of Biosearch Technologies, Inc.

³ HEX™ (Hexachloro-fluorescein) is a trademark of Applera Corp

⁴ TEX615™ is a trademark of Thermo Fisher Scientific.

⁵ BHQ-2™ (Black Hole Quencher) is a trademark of Biosearch Technologies, Inc.

⁶ Cy5™ is a trademark of GE Healthcare.

⁷ “TaqMan” is a trademark of Roche Molecular Systems, Inc.




Figure 1: Verification experiments with single targets (given in text boxes for each panel). All sets of probes and primers are present in every reaction, but positive signal is only observed when the target(s) is present, indicating that the amplification is specific.




Figure 2: Serial dilution experiments show LOD <10 molecules for the synthetic DNA construct of each target.

CONTACT US

For assistance, please contact DNA Software using the link: <https://www.pcrassays.com/contact/>

Address: Michigan Life Science and Innovation Center,
46701 Commerce Center Dr, Plymouth, MI 48170

Phone: (734) 222-9080